

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

Date of Report: 04/28/05

Date Received: 04/14/05

Project: Metro Self Monitor, PO# M95866, F&BI 504127

Date Extracted: 04/21/05

Date Analyzed: 04/21/05

**RESULTS FROM THE ANALYSIS OF THE WATER SAMPLE  
FOR TOTAL METALS  
BY EPA METHOD 6010**

Results Reported as mg/L (ppm)

<u>Sample ID</u> Laboratory ID	<u>Chromium</u>	<u>Copper</u>	<u>Nickel</u>	<u>Zinc</u>
M95866 504127-01	0.66	0.59	0.65	<0.05
Method Blank	<0.05	<0.05	<0.10	<0.05

# FRIEDMAN & BRUYA, INC.

## ENVIRONMENTAL CHEMISTS

Date of Report: 04/28/05

Date Received: 04/14/05

Project: Metro Self Monitor, PO# M95866, F&BI 504127

### QUALITY ASSURANCE RESULTS FOR TOTAL METALS BY INDUCTIVELY COUPLED PLASMA (ICP) (METHOD 6010)

Laboratory Code: 504135-05 (Duplicate)

Analyte	Reporting Units	Sample Result	Duplicate Result	Relative Percent Difference	Acceptance Criteria
Chromium	µg/ml (ppm)	<0.05	<0.05	nm	0-20
Copper	µg/ml (ppm)	<0.05	<0.05	nm	0-20
Nickel	µg/ml (ppm)	<0.1	<0.1	nm	0-20
Zinc	µg/ml (ppm)	<0.05	<0.05	nm	0-20

Laboratory Code: 504135-05 (Matrix Spike)

Analyte	Reporting Units	Spike Level	Sample Result	Percent Recovery MS	Acceptance Criteria
Chromium	µg/ml (ppm)	5	<0.05	99	50-150
Copper	µg/ml (ppm)	5	<0.05	96	50-150
Nickel	µg/ml (ppm)	10	<0.1	99	50-150
Zinc	µg/ml (ppm)	5	<0.05	100	50-150

Laboratory Code: Laboratory Control Sample

Analyte	Reporting Units	Spike Level	Percent Recovery LCS	Acceptance Criteria
Chromium	µg/ml (ppm)	5	95	70-130
Copper	µg/ml (ppm)	5	95	70-130
Nickel	µg/ml (ppm)	10	98	70-130
Zinc	µg/ml (ppm)	5	93	70-130

nm - The analyte was not detected in one or more of the duplicate analyses. Therefore, calculation of the RPD is not applicable.

504127

## SAMPLE CHAIN OF CUSTODY

CM  
ME 04-14-05

A15

Send Report To

G. Thompson

Company

Alaskan Copper Works

Address

628 S. Hancock St

City, State, ZIP

Seattle WA 98134

Phone #

206-782-8379

Fax #

206-382-4305

SAMPLERS (signature)

PROJECT NAME/NO.

Metro Self monitor

PO #

M95866

REMARKS

Page # \_\_\_\_\_ of \_\_\_\_\_

## TURNAROUND TIME

☐ Standard (2 Weeks)☒ RUSH 7 days

Rush charges authorized by: \_\_\_\_\_

## SAMPLE DISPOSAL

☐ Dispose after 30 days☐ Return samples☐ Will call with instructions

Sample ID	Lab ID	Date	Time	Sample Type	# of containers	ANALYSES REQUESTED										Notes
						TPH-Diesel	TPH-Gasoline	BTEX by 8021B	VOCs by 8260	SVOCs by 8270	HFS					
M95866	01	4/14	12:30	H2O	1											

Friedman & Bruya, Inc.  
3012 16th Avenue West

Seattle, WA 98119-2029

Ph. (206) 285-8282

Fax (206) 283-5044

FORMS\COC\COC.DOC

SIGNATURE

Relinquished by:

Received by:

Relinquished by:

Received by:

PRINT NAME

G. Thompson

Nhan Phan

COMPANY

Kew

FBI

DATE

4/14/05

4/14/05

TIME

1:03 pm

1:03

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

James E. Bruya, Ph.D.  
Charlene Morrow, M.S.  
Yelena Aravkina, M.S.  
Bradley T. Benson, B.S.  
Kurt Johnson, B.S.

3012 16th Avenue West  
Seattle, WA 98119-2029  
TEL: (206) 285-8282  
FAX: (206) 283-5044  
e-mail: fbi@isomedia.com

April 28, 2005

Gerry Thompson, Project Manager  
Alaskan Copper Works  
628 South Hanford  
Seattle, WA 98134


Dear Mr. Thompson:

Included are the results from the testing of material submitted on April 14, 2005 from the Metro Self Monitor, PO# M95866, F&BI 504127 project. There are 2 pages included in this report. Any samples that may remain are currently scheduled for disposal in 30 days. If you would like us to return your samples or arrange for long term storage at our offices, please contact us as soon as possible.

We appreciate this opportunity to be of service to you and hope you will call if you should have any questions.

Sincerely,

FRIEDMAN & BRUYA, INC.



Charlene Morrow  
Chemist

Enclosures  
ACU0428R.DOC



**FRIEDMAN & BRUYA, INC.**

**ENVIRONMENTAL CHEMISTS**

James E. Bruya, Ph.D.  
Charlene Morrow, M.S.  
Yelena Aravkina, M.S.  
Bradley T. Benson, B.S.  
Kurt Johnson, B.S.

3012 16th Avenue West  
Seattle, WA 98119-2029  
TEL: (206) 285-8282  
FAX: (206) 283-5044  
e-mail: fbi@isomedia.com

April 28, 2005

**DUPLICATE COPY**

**INVOICE # 05ACU0428-1**

Accounts Payable  
Alaskan Copper Works  
628 South Hanford  
Seattle, WA 98134

RE: Project Metro Self Monitor, PO# M95866, F&BI 504127 - Results of testing  
requested by Gerry Thompson for material submitted on April 14, 2005.

1 sample analyzed for Total Chromium, Copper, Nickel and Zinc  
by Method 6010B @ \$80 per sample \$ 80.00

Amount Due ..... \$ 80.00

FEDERAL TAX ID # (b) (6)